Radcal Corporation

426 W. Duarte Rd. Monrovia, CA 91016 Tel: (626) 357-7921

Fax: (626) 357-8863

Report No: S77586

Certificate of Calibration

Issued To: R-Tech Dental of Minnesota 1653 Bear Meadow Ln SE Rochester, MN 55904

Equipment Description Model S/N **Asset No.**Dent 50-100 kVp Test-O-Meter 4075D 80685 N/A

| Condition | of F | Equipment | as | Receiv | red |
|-----------|------|--------------|----|---------|-----|
| Condition | UI L | _uuiviiiciii | aэ | 1700014 | Cu |

In Calibration as Received: No

Remarks:

Upon receipt, the equipment was found to be in good condition and performed in accordance with the manufacturer's specifications with the exceptions listed below:

- 1. The 9V Battery measured less than 9V and was replaced in accordance with Radcal procedures.
- 2. The maximum error was found to be 3%. This exceeds the combined manufacturer's error (\pm 2%) and Radcal's statement of measurement accuracy (\pm 1%).
- 3. During kVp calibration the instrument displayed "Sh.E" at 99 kVp.

The equipment identified above has been calibrated and tested using standard Radcal calibration and acceptance procedures in accordance with Radcal Quality Manual PP1007 and technical requirements contained in the customer's order. These procedures are designed to ensure that the tested equipment meets or exceeds the stated specifications and the requirements of ANSI/NCLS Z540-1-1994.

All measurements performed during the testing employ equipment traceable to NIST or another recognized National Laboratory such as Physikalisch-Technische Bundesansalt (PTB). All calibration results included with this certificate were recorded at the time of measurement and shall not imply anything about the instrument's future stability. This must be determined from previous historical data.

The calibration results, any special limitations of use, environmental conditions during calibration, details of any maintenance such as service, adjustments, repairs or modifications are documented and found in the CALIBRATION REPORT.

Certificate Issue Date: September 30, 2004

P. Sunde / T. Harrington Authorized Reviewers

This Certificate of Calibration shall not be reproduced except in full, without the written approval of Radcal Corporation.

S77586 Dent1.xls 2/16/98 Page 1 of 3

Radcal Corporation

426 W. Duarte Rd. Monrovia, CA 91016 Tel: (626) 357-7921

Fax: (626) 357-8863

Report No: S77586

Date: September 30, 2004

Calibration Report

Measurement Test Conditions

An Electromed EDEC-80 X-ray generator equipped with an Eimac Model A192 tungsten target x-ray tube was used as the source of the required x-ray beam. The generator ripple is less than 0.5 kV. The inherent filtration was adjusted by the addition of aluminum to produce a total filtration of 2.5 mmAl in accordance with IPSM Report No, 64 "Data for Estimating X-ray Tube Total Filtration". The output of the generator was measured by a Radcal Dynalyzer IIIU. The Dynalyzer were monitored simultaneously by an analog to digital converter with an accuracy of \pm 0.1%. All reported measurement results have an accuracy of better than \pm 1% at the 95% confidence level.

Measurement Results

Conditions of Measurement

Temperature: 23 °C Pressure: 99.3 kPa Humidity: 39%

Note: Corrections for environmental conditions are not required for this equipment.

| Battery Check | Nom. (V) | Meas.(V) | |
|---------------|----------|----------|--|
| As Received | 9 | 8.85 | |
| As Left | 9 | 9.49 | |

kVp Calibration Calibration

2.5 mmAl total filtration, 22 inches target to detector

| | | | | Test-O-Meter | | |
|-----------|-------|------|------|--------------|------|-------|
| Dynalyzer | | | Read | Error | | |
| ms | mA | mAs | kVp | kVp | kVp | % |
| 254.1 | 192.1 | 48.8 | 51.6 | 51.5 | -0.1 | -0.2% |
| 254.8 | 191.7 | 48.8 | 54.6 | 54.3 | -0.3 | -0.6% |
| 253.8 | 192.7 | 48.9 | 59.8 | 60.2 | 0.4 | 0.6% |
| 257.2 | 190.0 | 48.9 | 70.1 | 70.6 | 0.5 | 0.7% |
| 258.5 | 188.9 | 48.8 | 80.4 | 82.9 | 2.5 | 3.0% |
| 256.3 | 190.6 | 48.9 | 89.7 | 91.4 | 1.7 | 1.9% |
| 254.2 | 192.5 | 48.9 | 94.8 | 96.6 | 1.8 | 1.9% |
| 253.9 | 192.4 | 48.9 | 99.0 | Sh.E | 0.0 | 0.0% |

Maximum Deviation:

3.0%

Pass/Fail (limit ±3%):

Pass

This Certificate of Calibration shall not be reproduced except in full, without the written approval of Radcal Corporation.

S77586 Dent1.xls 2/16/98 Page 2 of 3

Radcal Corporation

426 W. Duarte Rd. Monrovia, CA 91016 Tel: (626) 357-7921

Fax: (626) 357-8863

Report No: S77586

Date: September 30, 2004

